

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

**ANALYSIS REPORT, CERTIFIED**

*Material:* Conductivity Standard 84  $\mu$ s/cm

*Code:* C1049      *Report no:* RP8368      *Lot no:* 800228

**SPECIFICATION(S)**

**OBTAINED**

Conductivity @ 25°C

84  $\mu$ mhos/cm

*Analyst:* AB

*Approved* 

*Date:* 3/10/2008

**METTLER TOLEDO**

