

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413  $\mu\text{s}/\text{cm}$

Code: C1050 Report no: RP8389 Lot no: 800328

### SPECIFICATION(S)

Conductivity @ 25°C

### OBTAINED

1415  $\mu\text{mhos}/\text{cm}$

Analyst: MA

Approved: 

Date: 4/3/2008

**METTLER TOLEDO**