

**Buffer Quality Control Certificate**

Address	1900 Polaris Parkway, Columbus, OH 43240-4035
Phone	(800) METTLER (800) 638-8537
Fax	(614) 438-4900

[www.mt.com](http://www.mt.com)**Nominal Specifications:**

pH Standard Solution:	7 ( $\pm$ 0.02 at 25 °C)
Order Number:	63056125, 63056124, 63056123, 63056122, 30258825 (C0725)

**Buffer Properties:**

Actual pH:	7.00
Lot Number Solution:	17D235001 (MR17A11002)
Buffer Substance:	Potassium Phosphate, CAS # 7778-77-0
Production Date:	April, 23, 2017
Expiry Date:	April, 23, 2019

**Traceability:**

For the calibration of the value of our buffer or electrolyte solutions, the potentiometric measuring instruments were calibrated using one or two different reference buffer or electrolyte solutions, based on primary reference materials of the NIST<sup>1</sup>.

The National Metrology Institutes compare their measurement capabilities periodically in the key comparisons within the "Mutual Recognition Arrangement of National Measurement Standards and of Calibration Certificates issued by National Metrology Institutes (CIPM-MRA)". The international comparability, harmonization and the equivalence of their measurement capabilities are therefore ensured.

This guarantees a flawless traceability of the certified pH buffer values to NIST along with our Quality Management System. The certification is done for each lot.

**Standards:**

NIST Standard solution value:	6.864	9.186
NIST SRM :	186I – 186II	187e

<sup>1</sup> National Institute of Standards and Technology, Gaithersburg, USA

Date Certificate Issued:	May, 01, 2017
Quality Control (Print/Sign) Name	x Alicia Bourdeau 