

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4036 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Buffer Solution, pH 4.00, +/- 0.01, Color Coded Red

Code: C0723 Report no: RP3106 Lot no: 411103

### SPECIFICATION(S)

### OBTAINED

pH @ 25°C (at the time of packaging)

4.00

This standard buffer solution is an aqueous formulation of Potassium Hydrogen Phthalate, ACS and a suitable preservative. It is standardized to a pH of 4.00 +/- 0.01 @ 25°C using meters calibrated with NIST SRM 185g Potassium Hydrogen Phthalate. Final release criteria include the attainment of duplicate assays that are within the specified tolerances on separate meters. All volumetric glassware used in the standardization process meets the requirements of A.S.T.M. Class A tolerances. Balances are routinely calibrated according to our documented calibration program using check weights that are NIST traceable.

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SM

Expiration Date: 11/3/2016

Document Author: HMB

Manufacture Date: 11/3/2014

Approved: E. Chambers

Date: 11/17/2014

**METTLER TOLEDO**

