

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84 μ S/cm

Code: C1049 *Report no:* RP1252 *Lot no:* 111102

SPECIFICATION(S)	OBTAINED
Conductivity @ 25°C	83.9 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Manufacture Date: 11/2/2011

Approved 

Expiration Date: 11/2/2013

Date: 11/3/2011

METTLER TOLEDO

