

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84  $\mu$ S/cm

Code: C1049 Report no: RP3259 Lot no: 15B130003

### SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

83.9  $\mu$ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SM

Manufacture Date: 2/13/2015

Document Author: NB

Expiration Date: 2/13/2017

Approved HMB

Date: 2/17/2015

**METTLER TOLEDO**