

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
internet www.ml.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84 μ S/cm

Code: C1049 Report no: RP1420 Lot no: 210501

SPECIFICATION(S)

Conductivity @ 25°C

OBTAINED

84.1 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Manufacture Date: 5/1/2012

Approved: AB

Expiration Date: 5/1/2014

Date: 5/7/2012

METTLER TOLEDO