

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 μ S/cm

Code: C1050 Report no: RP1232 Lot no: 111012

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1414 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Expiration Date: 10/12/2013

Analyst: SM

Manufacture Date: 10/12/2011

Approved 

Date: 10/14/2011

METTLER TOLEDO