

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 μ S/cm

Code: C1050 *Report no:* RP3301 *Lot no:* 15C130016

SPECIFICATION(S)	OBTAINED
Conductivity @ 25°C	1412 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: TH

Expiration Date: 3/13/2017

Document Author: NB

Manufacture Date: 3/13/2015

Approved: EC

Date: 4/20/2015

METTLER TOLEDO