

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4036 USA
Phone 1-800-METTLER (1-800-638-6537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 $\mu\text{S}/\text{cm}$

Code: C1050 Report no: RP1640 Lot no: 211117

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1414 $\mu\text{S}/\text{cm}$

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Expiration Date: 11/17/2014

Document Author: NB

Manufacture Date: 11/17/2012

Approved: 

Date: 11/27/2012

METTLER TOLEDO

