

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4036 USA  
Phone 1-800-METTLER (1-800-638-8637)  
Fax 1-614-438-4625  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413  $\mu$ S/cm

Code: C1050 Report no: RP2808 Lot no: 410714

SPECIFICATION(S) **OBTAINED**

Conductivity @ 25°C 1412  $\mu$ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SM

Expiration Date: 7/14/2016

Document Author: NB

Manufacture Date: 7/14/2014

Approved: 

Date: 7/15/2014

**METTLER TOLEDO**

