

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 12.88 mS/cm

Code: C1051 Report no: RP1228 Lot no: 111005

SPECIFICATION(S)

Conductivity @ 25°C

OBTAINED

12.89 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SM

Manufacture Date: 10/5/2011

Approved: 

Date: 10/6/2011

Expiration Date: 10/5/2013

METTLER TOLEDO