

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4036 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 12.88 mS/cm

Code: C1051 Report no: RP3322 Lot no: 15D130002

SPECIFICATION(S)

Conductivity @ 25°C

OBTAINED

12.86 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SM

Expiration Date 4/13/2017

Document Author: RAS

Manufacture Date: 4/13/2015

Approved: AB

Date: 5/12/2015

METTLER TOLEDO