

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

**ANALYSIS REPORT, CERTIFIED**

Material: Conductivity Standard 12.88 mS/cm

Code: C1051 Report no: RP1330 Lot no: 210228

**SPECIFICATION(S)**

**OBTAINED**

Conductivity @ 25°C

12.86 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Manufacture Date: 2/28/2012

Approved: AB

Date: 3/2/2012

Expiration Date: 2/28/2014

**METTLER TOLEDO**