

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: KCl Solution 3 mol/L

Code: C4190 *Report no.:* RP3229 *Lot no.:* 510114

SPECIFICATION(S)

OBTAINED

Assay

3.07 M

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Expiration Date: 1/14/2017

Document Author: NB

Manufacture Date: 1/14/2015

Approved: 

Date: 1/19/2015

METTLER TOLEDO

