

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Silver Nitrate Solution, 0.05 Normal

Product Code: SN005 Report no: RP1464 Lot no: 210621

**SPECIFICATION(S)** **OBTAINED**

Assay 0.0501N

This standard volumetric solution is prepared using the highest quality reagent grade raw materials available. It is standardized against Potassium Chloride, ACS that is directly traceable to NIST SRM 999 Potassium Chloride Argentometric Standard. The assay procedure employed is in accordance with current industry standard protocols. Final release criteria include the attainment of duplicate assays that are within the specified tolerances. All volumetric glassware used in the standardization process meets the requirements of A.S.T.M. Class A tolerances. Balances are routinely calibrated according to our documented calibration program using check weights that are NIST traceable.

Analyst: SM

Manufacture Date: 6/21/2012

Approved: AB

Expiration Date: 6/21/2013

Date: 6/25/2012

**METTLER TOLEDO**