

# Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone | 1-800-METTLER (1-800-638-8537)  
Fax | 1-614-438-4525  
Internet | www.mt.com

## ANALYSIS REPORT, CERTIFIED

*Material:* Conductivity Standard 84  $\mu$ s/cm

*Code:* C1049      *Report no:* RP8543      *Lot no:* 801028

### SPECIFICATION(S)

Conductivity @ 25°C

### OBTAINED

84.1  $\mu$ mhos/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

*Analyst:* SB

*Manufacture Date:* 10/28/2008

*Approved* NB

*Date:* 10/28/2008

**METTLER TOLEDO**