

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84 μ s/cm

Code: C1049 *Report no:* RP8642 *Lot no:* 900324

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

84.2 μ mhos/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SNB

Manufacture Date: 3/24/2009

Approved NB

Date: 3/31/2009

