

# Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone | 1-800-METTLER (1-800-638-8537)  
Fax | 1 614 438-4525  
Internet | www.mt.com

## ANALYSIS REPORT, CERTIFIED

*Material:* Conductivity Standard 84  $\mu$ s/cm  
*Code:* C1049    *Report no.:* RP8794    *Lot no.:* 900928

### SPECIFICATION(S)

Conductivity @ 25°C

### OBTAINED

84.0  $\mu$ mhos/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

*Analyst:* AB

*Manufacture Date:* 9/28/2009

*Approved* NB

*Date:* 9/29/2009

**METTLER TOLEDO**