

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 12.88 mS/cm
Code: C1051 *Report no:* RP8391 *Lot no:* 800108

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

12.87 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: AB

Manufacture Date: 1/8/2009

Approved MA

Date: 4/3/2008

METTLER TOLEDO