

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

*Material:* Conductivity Standard 12.88 mS/cm

*Code:* C1051    *Report no:* RP8679    *Lot no:* 801120

### SPECIFICATION(S)

Conductivity @ 25°C

### OBTAINED

12.87 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

*Analyst:* AB

*Manufacture Date:* 11/20/2008

*Approved* NB

*Date:* 5/19/2009

**METTLER TOLEDO**