

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413  $\mu\text{s/cm}$

Code: C1050 Report no: RP8834 Lot no: 901119

### SPECIFICATION(S)

### OBTAINED

Conductivity @ 25°C

1413  $\mu\text{S/cm}$

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: AB

Manufacture Date: 11/19/2009

Approved M.A.

Date: 11/20/2009

**METTLER TOLEDO**