

# Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone | 1-800-METTLER (1-800-638-8537)  
Fax | 1-614-438-4525  
Internet | www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84  $\mu$ S/cm

Code: C1049 Report no: RP8847 Lot no: 901222

### SPECIFICATION(S)

### OBTAINED

Conductivity @ 25°C

83.9  $\mu$ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: AB

Manufacture Date: 12/22/2009

Approved: M.A.

Date: 12/23/2009

**METTLER TOLEDO**