

Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone | 1-800-METTLER (1-800-638-8637)
Fax | 1-614-438-4525
Internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 $\mu\text{S}/\text{cm}$

Code: C1050 Report no: RP8844 Lot no: 901217

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1412 $\mu\text{S}/\text{cm}$

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: AB

Manufacture Date: 12/17/2009

Approved M.A.

Date: 12/21/2009

METTLER TOLEDO