



Anachemia

Anachemia Chemicals Inc.

3 Lincoln Blvd.
Rouses Point, NY
12979
Tel.: (518) 297-4444
Fax: (518) 297-2960
www.anachemiachemicals.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 $\mu\text{s}/\text{cm}$

Code: C1050 Report no: RP8975 Lot no: 010624

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1412 $\mu\text{S}/\text{cm}$

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Expiration Date: 6/24/2012

Analyst: AB

Manufacture Date: 6/24/2010

Approved M.A.

Date: 6/25/2010