

Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway Columbus OH 43240-4035 USA
Phone | 1-800-METTLER (1-800-638-8537)
Fax | 1-614-438-4525
Internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 μ S/cm

Code: C1050 *Report no.:* RP9000 *Lot no.:* 010729

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1411 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Expiration Date: 7/29/2012

Analyst: AB

Manufacture Date: 7/29/2010

Approved 

Date: 8/2/2010

