

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84 μ S/cm

Code: C1049 Report no: RP9059 Lot no: 011103

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

84.2 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: AB

Manufacture Date: 11/3/2010

Approved NB

Date: 11/4/2010

Expiration Date: 11/4/2012

METTLER TOLEDO