

# Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone 1-800-METTLER (1-800-638-8537)  
Fax 1-614-438-4525  
Internet www.mt.com

## ANALYSIS REPORT, CERTIFIED

*Material:* Conductivity Standard 1413  $\mu\text{S}/\text{cm}$

*Code:* C1050      *Report no.:* RP1127      *Lot no.:* 110615

### SPECIFICATION(S)

**OBTAINED**

Conductivity @ 25°C

1414  $\mu\text{S}/\text{cm}$

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

*Analyst:* AB

*Expiration Date:* 6/15/2013

*Manufacture Date:* 6/15/2011

*Approved* NB

*Date:* 6/20/2011

**METTLER TOLEDO**