

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone | 1-800-METTLER (1-800-638-8537)
Fax | 1-614-438-4525
Internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 1413 μ S/cm

Code: C1050 *Report no:* RP1193 *Lot no:* 110817

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

1414 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Expiration Date: 8/17/2013

Analyst: EC

Manufacture Date: 8/17/2011

Approved SM

Date: 8/19/2011

METTLER TOLEDO