

Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway Columbus OH 43240 4035 USA
Phone | 1-800-METTLER (1-800-638-8537)
Fax | 1-614-438-4525
internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 12.88 mS/cm
Code: C1051 Report no: RP1180 Lot no: 110802

SPECIFICATION(S)

OBTAINED

Conductivity @ 25°C

12.89 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Manufacture Date: 8/2/2011

Approved: AB

Date: 8/3/2011

Expiration Date: 8/2/2013

