

# Mettler-Toledo, Inc.

Address: 1900 Polaris Parkway, Columbus, OH 43240-4035 USA  
Phone: 1-800-METTLER (1-800-638-8537)  
Fax: 1-614-438-4525  
Internet: www.mt.com

## ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 12.88 mS/cm  
Code: C1051 Report no: RP1210 Lot no: 110909

### SPECIFICATION(S)

Conductivity @ 25°C

**OBTAINED**

12.89 mS/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC  
Approved: AB

Manufacture Date: 9/9/2011

Date: 9/12/2011

Expiration Date: 9/9/2013

**METTLER TOLEDO**