

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone | 1-800-METTLER (1-800-638-8537)
Fax | 1-614 438-4525
Internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Potassium Biphthalate, Crystals

Code: 7576 *Report no:* RP4002 *Lot no:* 17G185006

SPECIFICATION(S)

OBTAINED

Assay (Dried basis: C8H5O4K)	100.00%
pH of 0.05m solution at 25 (+-) 0.2°C	4.01
Insoluble matter	<0.005%
Chlorine compounds (as Cl)	<0.003%
Sulfur compounds (as S)	<0.002%
Heavy metals (as Pb)	<5 ppm
Iron (Fe)	<5 ppm
Sodium (Na)	<0.005%

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: SA

Document Author: NB

Approved: AB
(On behalf of Mettler Toledo)

Expiration Date: 18 Jul 2019

Manufacture Date: 18 Jul 2017

Date: 02 Aug 2017

METTLER TOLEDO