

Mettler-Toledo, Inc.

Address | 1900 Polaris Parkway, Columbus, OH 43240-4036 USA
Phone | 1-800-METTLER (1-800-638-8537)
Fax | 1-614-438-4525
Internet | www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Buffer Solution, pH 10.00, +/- 0.01, Color Coded Blue

Code: C0728 Report no: RP1548 Lot no: 211003

SPECIFICATION(S)

OBTAINED

pH @ 25°C (at the time of packaging)

10.01

This standard buffer solution is an aqueous formulation of ACS grade Potassium Chloride, Boric Acid, Sodium Hydroxide and a suitable preservative. It is standardized to a pH of 10.00 +/- 0.01 @ 25°C using meters calibrated with NIST SRMs 186 Potassium Dihydrogen Phosphate/186 Disodium Hydrogen Phosphate and 187 Sodium Tetraborate decahydrate Ion Activity Standards. Final release criteria include the attainment of duplicate assays that are within the specified tolerances on separate meters. All volumetric glassware used in the standardization process meets the requirements of A.S.T.M. Class A tolerances. Balances are routinely calibrated according to our documented calibration program using check weights that are NIST traceable.

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: EC

Expiration Date 10/3/2014

Approved NB

Manufacture Date: 10/3/2012

Date: 10/5/2012

METTLER TOLEDO

