

Buffer Quality Control Certificate

Address | 1900 Polaris Parkway, Columbus, OH 43240-4035
 Phone | (800) METTLER
 | (800) 638-8537
 Fax | (614) 438-4900

www.ml.com**Nominal Specifications:**

pH Standard Solution:	5 (± 0.02 at 25 °C)
Order Number:	30258823, 30258824 (C0772)

Buffer Properties:

Actual pH:	5.00
Lot Number Solution:	16K155002
Buffer Substance:	Potassium Biphthalate, CAS # 877-24-7
Production Date:	November, 15, 2016
Expiry Date:	November, 15, 2018

Traceability:

For the calibration of the value of our buffer or electrolyte solutions, the potentiometric measuring instruments were calibrated using one or two different reference buffer or electrolyte solutions, based on primary reference materials of the NIST¹.

The National Metrology Institutes compare their measurement capabilities periodically in the key comparisons within the "Mutual Recognition Arrangement of National Measurement Standards and of Calibration Certificates issued by National Metrology Institutes (CIPM-MRA)". The international comparability, harmonization and the equivalence of their measurement capabilities are therefore ensured.

This guarantees a flawless traceability of the certified pH buffer values to NIST along with our Quality Management System. The certification is done for each lot.

Standards:

NIST Standard solution value:	4.005 \pm 0.005 (25°C)	6.864 \pm 0.005 (25°C)
NIST SRM:	185i	186I – 186II

¹ National Institute of Standards and Technology, Gaithersburg, USA

Date Certificate Issued:	May, 01, 2017
Quality Control (Print/Sign) Name	x Alicia Bourdeau/ 