

Mettler-Toledo, Inc.

Address 1900 Polaris Parkway, Columbus, OH 43240-4035 USA
Phone 1-800-METTLER (1-800-638-8537)
Fax 1-614-438-4525
Internet www.mt.com

ANALYSIS REPORT, CERTIFIED

Material: Conductivity Standard 84 μ S/cm
Code: C1049 **Report no:** RP4010 **Lot no:** 17H025004

SPECIFICATION(S)	OBTAINED
Conductivity @ 25°C	83.8 μ S/cm

Products will meet specifications outlined in this C of A for a period of up to 2 years from the date of manufacture stored unopened at room temperature.

Analyst: <u>SM</u>	Manufacture Date: <u>8/2/2017</u>
Document Author: <u>NB</u>	Expiration Date: <u>8/2/2019</u>
Approved: <u></u> (Signed on behalf of Mettler Toledo)	Date: <u>8/3/2017</u>

METTLER TOLEDO